

HAZARD MANAGEMENT SAFE OPERATING PROCEDURE (SOP)

Only to be completed where required as a control measure under a Risk Assessment

NAME OF THE TASK/ACTIVITY	NIKON LIVE CELL MICROSCOPE	DATE	24/02/2020
LOCATION	ADELAIDE MICROSCOPES BUILDING RM 1.22	Insert photo (Optional)	
Hazards identified on the RA	Contact with electricity or potential for electric shock Exposure to non-ionising radiation (UV light) Mercury lamp breakage and vapour or liquid leak		
PERSONAL PROTECTIVE EQUIPMENT			
Eye protection: Safety glasses • Eye shields • Safety goggles • Other: Hand protection: Rubber • Cut resistant • Leather • Vinyl • Neoprene Nitrile • Barrier creams •			

HSW Handbook

2 Eon... (h)TJ ET 0 0.00 6.938 763 66d61400 6d

HSW Handbook	Hazard Management	Effective Date:	17 December 2019	Version 3.0
Authorised by	Chief Operating Officer (University Operations)	Review Date:	17 December 2022	Page 2 of 2
Warning	This process is uncontrolled when printed. The current version of this document			